

All-India Institute of Medical Sciences
Ansari Nagar, New Delhi-29
(RESEARCH SECTION)

Ref. No.24/Prop/Plastic Surg./RT/21-22/RS

Dated: 15.3.2022

Subject: Procurement of Venous Anastomotic Coupler (Disposable) on proprietary basis - Inviting comments thereon.

The request has been received from **Raja Tiwari** to purchase the subject item from M/s. Myovatec Surgical System Pvt. Ltd on proprietary basis. The proposals submitted by M/s. Myovatec Surgical System Pvt. Ltd. and Performa Invoice and Departmental PAC certifications are attached.

The above documents are being uploaded for open information to submit objections, comments, if any, from any manufacturer regarding proprietary nature of the equipment/item within issue of 15 days giving reference **No.24/Prop/Plastic Surg./RT/21-22/RS**. The comments should be received by office of Stores Officer (RS), Research Section at AIIMS on or before 31.3.2022 upto 12:00 P.M., failing which it will be presumed that any other vendor is having no comment to offer and case will be decided on merits.


ASST. STORES OFFICER (RS)

Encl: Related documents enclosed.

1. PAC Certificate enclosed.
2. Performa Invoice

Myovatec Surgical Systems Pvt. Ltd.

A-12/42A, M.I.G. D.D.A. FLATS, KALKAJI EXTENSION, NEW DELHI-110019. INDIA

Phone : 29985846, 29980759 Fax : 91-11-29981451

E-mail : myovatec@gmail.com reception.myovatec@gmail.com



An ISO 9001 Company

To,
Dr. Raja Tiwari,
Professor,
Deptt. of Plastic Surgery,
AIIMS,
New Delhi

Quote:266

Date:13/11/21

S.No.	DESCRIPTION	QTY	Rate (Rs.)	Amount (Rs.)
	(Mfd By Synovis,USA)			
1	Gem Coupler	11 Pcs	16,000.00	1,76,000.00
			Total	1,76,000.00
			GST 12%	21,120.00
			G.Total	1,97,120.00

Terms & Conditions

- a Warranty & CMC not applicable since stores are consumable item
- b CMC Not applicable since store is a consumable item
- c List of all spares, accessories, consumables etc with price should be quoted separately for future purchases.(No accessories needed)
- d Standard payment terms as per AIIMS purchase manual.
- 12 Copy of latest valid authorization certificate Attached

For Myovatec Surgical Systems Pvt. Ltd.

For Myovatec Surgical Systems Pvt. Ltd.

[Signature]
Director

BANK DETAILS:
Bank Name : I C I C I Bank Limited
Branch : D-138, East of Kailash,
New Delhi - 110019
A/c No. : '071905000334
MICR Code : '110229080
RTGS/NEFT/IFSC CODE : ICIC0000719
A/c Name : Myovatec Surgical Systems Pvt. Ltd.
Address : A-12/42A, Kalkaji Extension
New Delhi - 110019



ATTESTED

NOTARY PUBLIC

31 JAN 2022

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI-110029

(6050)
11

RESEARCH SECTION

PROPRIETARY/SPECIFIC BRAND GOODS CERTIFICATE

1. Item/Type/Model No. required along with specification. *Anatomotic coupler (various sizes).*
2. Is the item a spare parts attachment or accessory for an existing equipment. *Yes, above couplers are consumable items of GEM COUPLING EQUIPMENT*
3. Name of the manufacturers/supplier of the item proposed by the Indentor. *GEM, Synovis*
4. Are they sole manufacturers/sole distributors of the item. *Yes*
5. Is there any other item with similar/ equivalent specification available in the market to meet the job requirement envisaged. If the answer is yes, why the same can't be procured. Demanding officer should bring out comparative functional advantages/cost effectiveness of the recommended item from these offered by other. *NO*
6. What were the efforts made to locate alternative source of supply or use other substitutes. *ONLINE X PHYSICAL Resources*
7. Why open/limited tender can't be resorted to, for locating alternative source. *Proprietary item*
8. Are the proprietary items certifying that the rates are reasonable or not. *Yes*
9. Any other justification for procuring item from single source. *No other company manufactures similar device. The device is already procured in department.*

Signature of Indentor *[Signature]*
(Demanding Officer) *Dr. Raja Tiwari*
Associate Professor
Burns & Plastic Surgery Block

COUNTERSIGNATURE *27/11/2021*
(Head of the Department) *[Signature]*
Prof. & Head
Burns & Plastic Surgery Block

I certify that the item above is required to be procured on single tender basis as the source of supply is definitely known and the brand proposed was advantages in meeting our functional requirements and limited tender system could be dispensed with as they would serve no useful purpose in this particular case.

(Strike out whichever is not applicable)

Research Section/Research Society
All India Institute of Medical Sciences, New Delhi
27 NOV 2021

DERMA-GRIP

Myovatec Surgical Systems Pvt. Ltd.

A-12/42A, M.I.G. D.D.A. FLATS, KALKAJI EXTENSION, NEW DELHI-110019. INDIA

Phone : 29985846, 29980759 Fax : 91-11-29981451

E-mail : myovatec@gmail.com reception.myovatec@gmail.com



An ISO 9001 Company

To
The HOD
Department of Plastic Surgery
AIIMS
New Delhi

Dated:-31/01/2022

Confirmation of supply for authorized agent

It is confirmed that the supplies will be executed by Myovatec Surgical Systems, Pvt. Ltd.

The exclusive authorized importer & distributor of OEM valid authority certificate from manufacturer is attached

Thanking you,
Sincerely yours,

For Myovatec Surgical Systems (P) Ltd.

For Myovatec Surgical Systems Pvt. Ltd.

(Authorised signatory)

Director



ATTESTED

NOTARY PUBLIC

31 JAN 2022



December 2, 2021

To,
Adesh Mathur
Myovatec Surgical Systems Pvt. Ltd.
A-12/42A, Kalkaji Extension
New Delhi – 110019
(India)

Dear Sir / Madam,

Equipment Name: Venous Coupler Device for Micro venous Anastomosis

Synovis Micro Companies Alliance, Inc., a wholly owned subsidiary of Baxter International, Inc., declare that we are the original manufacturer of the above described product (the GEM Coupler System and Coupler implants). Synovis Micro Companies Alliance, Inc., is located at 439 Industrial Lane, Birmingham, Alabama, 35211, U.S.A., phone number +1.205.941.0111, fax number +1.205.941.1522. Director of Sales is Terry L. Harrell, terry_harrell@baxter.com.

Myovatec Surgical Systems Pvt. Ltd., is our authorized, exclusive distributor for the country of India. The authorization is valid for a period of five years through 2026.

Name: Terry L. Harrell

Title: Director of Sales

Date: December 2, 2021

Signature: 
D4FFB2A4219B48B...



ATTESTED PHOTO COPY


Notary Public, Delhi

March 2, 2021


TO WHOM IT MAY CONCERN:

Effective as of 02 March, 2021, Synovis Micro Companies Alliance, Inc, a wholly owned subsidiary of Baxter International, Inc., at 439 Industrial Lane, Birmingham, Alabama 35211, USA, declare that Myovatec Surgical Systems Pvt. Ltd., an Indian Corporation and their designated agents, are our esteemed and exclusive distributors for the below listed products for the country of India:

- GEM COUPLER System
- GEM FLOWCOUPLER System
- GEM MicroClips and Appliers
- GEM Non Stick Bipolar Forceps

Thank you for your interest in the above products used in reconstructive microsurgical procedures.


Best Regards,



Terry L. Harrell
Director of Sales
Synovis Micro Companies Alliance

Subscribed and sworn before me this 2nd day of March, 2021.

Notary Signature



3/2/2021



TRACIE W. MAIZE
Notary Public, State of Alabama
Alabama State At Large
My Commission Expires
July 13, 2022



Micro Companies Alliance, Inc.


9

02 July 2021

Proprietary Certificate

We certify that the GEM™ Microvascular Anastomotic COUPLER™ manufactured of High density polyethylene (HDPE) rings, with stainless steel pins, available in variable diameters, is the proprietary product of *Synovis Micro Companies Alliance, Inc., a division Synovis Life Technologies, Inc., a subsidiary of Baxter International Inc., 439 Industrial Lane, Birmingham, Alabama, 35211, U.S.A., and is exclusively imported and distributed by Myovatec Surgical Systems Pvt. Ltd., A-12/8A, Kalkaji Extension, New Delhi 110019, India.*

As a responsible representative of *Synovis Micro Companies Alliance, Inc., a division Synovis Life Technologies, Inc., a subsidiary of Baxter International Inc.,* I hereby declare that the information contained within the proprietary certificate is truthful and accurate.


Julie S. Carlston
Senior Manager Regulatory Affairs

07-02-2021
Date

Technical specifications for Venous Anastomotic Coupler
(Disposable)- 10 (quantity)

1. The disposable venous couplers should be able to load on Anastomotic instrument.
2. These should be made of polyethylene rings with stainless steel pins.
3. The rings should be loaded on the inserter applicator plastic unit
4. The rings should be of different vessel sizes ranging from 1.0mm,1.5mm,2.0mm,2.5mm,3.0mm,3.5mm each size ring
5. Should have colour coding for easy identification
6. The s.s. pins in each ring Should be arranged in concentric manner & each pin should look into opposite ring pocket to perform complete coupling .