

**OFFICE OF THE DIRECTOR
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI-29
(DEPARTMENT OF PHARMACOLGY)**

R.E.No.7/Pharma/14-15

Dated 17/01/2015

Sealed quotations are invited by post/per bearer for the supply of items detailed below which are required for Department of Pharmacology, AIIMS, New Delhi-110029.

The quotations should reach this office on or before 05-02-2015 at 11.00 AM duly sealed & marked Tender/R.E. No. 7/Pharma/14-15. All quotations should be printed. Over writing and erased entries will not be considered and treated as deleted entries from the quotations. The validity of the quotation should be for six months.

The tenders are to be deposited with bid security of Rs. 5,000/= (Refundable) (Rupees Five thousand only) through a Demand Draft/Pay order drawn in favour of the "Director, AIIMS, New Delhi", along with the quotation failing which the quotation will not be considered. In case of failure to comply with the supply order in terms of the contract, the order will be placed to the next successful bidder and the bid security money deposited will be forfeited. The approved firm has to supply the item/material within a period of 30 days after issue of confirmed supply order.

The firm has to quote with literature, catalogue (if available).

The quotations/tenders will be opened on 05-02-2015 at 11.30 AM in the Room No 4013, Department of Pharmacology, AIIMS, New Delhi- 110029 in the presence of the tenderers, whosoever are present.

Sr. No. Nomenclature

Qty. Req.

1. Laboratory Hot Air Oven 2 nos.

Specification for Laboratory Hot Air Oven

1. Temperature range : 50⁰C to 250⁰C (accuracy $\pm 2^0$ C)
2. Oven should be corrosion resistant.
3. Double walled corrosion inner chamber made of stainless steel
4. Gap between two walls field with glass wool insulation to avoid heat loss
5. Fully insulated door with lock hinges and ventilation
6. Number of Shelves (3-4). Trays should be with GI Wire Mesh
7. Temperature should be thermostat controlled by electronic digital temperature controller with sensor (control switches)
8. Electronic digital timer Temperature display
9. Warranty : Two years

Dr. D.S.Arya
Prof. & O/I (Store)

Dr. Y.K.Gupta
Prof. & Head

Note:- Sample of the quoted item must be submitted along with quotation failing which offer will not be considered for acceptance.