



COMPREHENSIVE RURAL HEALTH SERVICES PROJECT
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
BALLABGARH - 121004 (HR)

Ph. 91- 0129-2241362
Fax. 91-0129-2211227

No.F.R.E. /CRHSP-AIIMS/Ballabgarh

Dated the: 01.07.2016

To,

Subject: Quotation for purchase of **Cautery Machine** at CRHSP, Ballabgarh Hospital.

Sir,

Your are requested to please give your quotation for purchase of **Cautery Machine** at CRHSP, Ballabgarh Hospital. Quotation should reach the office of the **undersigned latest by 15.07.2016**. Specification is as under:-

- | | | |
|--------------------------|---|---------------------------------|
| • Output frequency | - | Monopolar generator 500±25KHz |
| • Cut Control (Pure) | - | 400 Watts 500 ohms load |
| • Cut Control (Blend-II) | - | 300 Watts 500 ohms load |
| • Cut control (Blend-I) | - | 225 Watts 500 ohms load |
| • Coag Control | - | 160 watts 500 ohms load |
| • Bipolar Coag | - | 100 Watts 125 ohms load |
| • Spray coag | - | 100 watts 125 ohms load |
| • Input voltage | - | 220 240 volts AC 50 Hz |
| • Power Consumption | - | 750 VA apporx. |
| • Dimensions | - | (39 cm W x 33.5cm D x 13.5cm H) |

Terms and Conditions:-

Comprehensive warranty for two years from the date of purchase.

Standard AIIMS terms and conditions apply.

The responsibility for warranty will be entirely with the local dealer for the full period.

Thanking you,

(Dr. Shashi Kant)
Professor-in-charge
CRHS Project, Ballabgarh