

Bid Document

Bid Details	
Bid End Date/Time	06-08-2022 12:00:00
Bid Opening Date/Time	06-08-2022 12:30:00
Bid Offer Validity (From End Date)	65 (Days)
Ministry/State Name	Ministry Of Health And Family Welfare
Department Name	Department Of Health And Family Welfare
Organisation Name	All India Institute Of Medical Sciences (aiims)
Office Name	Aiims, New Delhi
Total Quantity	400
Item Category	Oxygenator Membrane with Hard Shell Reservoir (Q3)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Bid to RA enabled	Yes
RA Qualification Rule	H1-Highest Priced Bid Elimination
Time allowed for Technical Clarifications during technical evaluation	2 Days
Estimated Bid Value	6500000
Evaluation Method	Total value wise evaluation

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

MII Purchase Preference

MII Purchase Preference	Yes
-------------------------	-----

MSE Purchase Preference

MSE Purchase Preference	Yes
-------------------------	-----

1. Preference to Make In India products (For bids < 200 Crore): Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .In case Buyer has selected Purchase preference to Micro and Small Enterprises clause in the bid, the same will get precedence over this clause.

2. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for 25%(selected by Buyer) percentage of total QUANTITY.

3. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

4. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:

- i. If number of technically qualified bidders are only 2 or 3.
- ii. If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
- iii. In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
- iv. If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1
- v. If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1

Oxygenator Membrane With Hard Shell Reservoir (400 pieces)

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

Brand Type	Registered Brand
------------	------------------

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Bid Requirement (Allowed Values)
PRODUCT INFORMATION	Size	Paediatric, Semi Adult, Adult, Neonatal
	Capacity of hard shell venous reservoir	600 ml or more, 1000 ml or more, 3000 ml or more, 4000 ml or more
	Blood flow range	Between 0.5 to 1.2 Ltr/min, Between 0.5 to 2.8 Ltr/min, Between 0.5 to 7.0 Ltr/min, Between 0.5 to 3.5 Ltr/min
	Priming volume	Less than 70 ml, Between 70 ml to 100 ml, Between 100 ml to 140 ml, Between 140 ml to 180 ml, Between 180 ml to 320 ml
	Venous return port	With rotating Facility, With rotating Facility and Reducer
	Size of venous return port	1/2 inch, 1/4 inch, 3/8 inch
	Arterial outlet port size	1/4 inch, 3/8 inch
	Venous outlet port size	1/2 inch, 1/4 inch, 3/8 inch
	Venous inlet port size	1/2 inch, 1/4 inch, 3/8 inch, 3/8 inch with reducers, 1/4 inch with reducers
SHELF LIFE	Shelf Life from the date of manufacture	12, 18, 24 Or higher

Additional Specification Parameters - Oxygenator Membrane With Hard Shell Reservoir (400 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Technical Specification1	Oxygenator Adult (Integrated with Arterial Filter & Heat Exchanger) 1. Oxygenator should have integrated arterial filter with cardiomy/venous reservoir 2. Should have integrated arterial filter with self venting technology 3. Heat exchanger surface area should be not more than 0.2m ² 4. Venous filter should <50 micron meter 5. Priming volume should not be more than 300ml
Technical Specification2	6. Blood Flow range should be 0.5 to 7LPM 7. Heat Exchanger efficiency should not be less than 0.50 at max flow 8. Pressure drop should be minimum upto 110mmgh or less 9. Arterial Filter should be <35 micron meter 10. Membrane surface area should be 2-2.5m ²
Advance Sample	The Selection will be made on the basis of evaluation of sample submitted by the bidder firm therefore, the firm must submit at least 01(one) advance sample for quality verification before the bid end date & time to CTVS Package Store, CT-4 Ward, 4th Floor, CN Centre, AIIMS, New Delhi-110029. The bid details (Bidder name, and Bid Number) should be clearly mentioned on each sample.

Specification Parameter Name	Bid Requirement (Allowed Values)
Unsuccessful Bids/Samples/Rejected Samples	The unqualified bidders should collect their samples within 07 days from the updation of sample/technical selection/rejection remarks on GeM AND The unsuccessful bidders whose samples are technically approved but could not get the order should collect their sample within 07 days after awarding of bid. Failing which the competent authority reserves the right to take appropriate decision for the uncollected samples and no representation will be entertained further in this regard.

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Delivery Schedule (In number of days from contract start days)		
			Quantity	Delivery to start after	Delivery to be completed by
1	Amit Kumar	110029,Ansari Nagar			
			150	15	30
			150	31	45
			100	61	75

Buyer Added Bid Specific Terms and Conditions

1. Make in india specific authorisation certificate needs to be enclosed.
2. **Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

3. **Buyer Added Bid Specific ATC**

Buyer Added text based ATC clauses

The Selection will be made on the basis of evaluation of sample submitted by the bidder firm therefore, the firm must submit at least 01(one) advance sample for quality verification before bid end date and time to CTVS Package Store, CT-4 Ward, 4th Floor, CN Centre, AIIMS, New Delhi-110029. The bid details (Bidder Name, and Bid Number) should be clearly mentioned on each sample. The unqualified bidders should collect their samples within 07 days from the updation of sample/technical selection/rejection remarks on GeM AND The unsuccessful bidders whose samples are technically approved but could not get the order should collect their sample within 07 days after awarding of bid. Failing which the competent authority reserves the right to take appropriate decision for the uncollected samples and no representation will be entertained further in this regard. In case of any ambiguity in technical specification, the technical specification mentioned as additional parameter will be considered as final and no representation in this regard will be entertained.

Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization. Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity/restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and/or terms and conditions governing the bid. Any clause incorporated by the Buyer such as demanding Tender Sample, incorporating any clause against the MSME policy and Preference to make in India Policy, mandating any Brand names or Foreign Certification, changing the default time period for Acceptance of material or payment timeline governed by OM of Department of Expenditure shall be null and void and would not be considered part of bid. Further any reference of conditions published on any external site or reference to external documents/clauses shall also be null and void. If any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations. Also, GeM does not permit collection of Tender fee / Auction fee in case of Bids / Forward Auction as the case may be. Any stipulation by the Buyer seeking payment of Tender Fee / Auction fee through ATC clauses would be treated as null and void.

[This Bid is also governed by the General Terms and Conditions](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---Thank You---