

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
DEPARTMENT OF ANATOMY
Ansari Nagar, New Delhi-110029

RE/ 2 /Anat/Store/16-17

Dated : 23/11/2016

Dear M/s,

A short rate enquiry is hereby floated/invited for the following item required by Department of Anatomy, AIIMS , New Delhi-110029.


1.	Plethysometer for rat and mouse
2.	Perfusion pump for rat
3.	Digital colour camera foe existing Nikon 80 i microscope
4.	Refrigerator 4 degree
5.	Analytical electronic weighing balance (Micrograms)
6.	Manual Rotary Microtome for cutting sections
7.	Pipette aid

Firm submit the following:

1. Firm should submit price Reasonable Certificate
2. Firm should submit the under taking on the letter that the firm has not been black listed
3. Firm should submit the under taking on the letter that the firm has no vigilance case pending against Supply /firm
4. Firm should submit all the certifications asked for in the technical specifications

Terms & Conditions:

- The quotations should be submitted to the Professor & Head, Department of Anatomy, Room No. 1010, AIIMS, New Delhi-29 and sent by post/per bearer **by 14th December, 2016 up to 12.00 P.M.** Quotation should be sealed in an envelope and the "**name of the item**" clearly written on top of the envelop.
 - The make of the article offered should invariably be quoted. Quotation should be typed written in ink. All overwriting and erased entries will be deleted from the quotation. The rates should be valid for at least THREE MONTHS.
 - VAT/Sales tax or any other kind of tax(s) must be separately mentioned against each item. In case no sales tax is chargeable, prices must be quoted as NET.
 - **Warranty/Guarantee:** two years followed by three years CMC comprehensive warranty.
 - **Installation:** The firm install the equipment on the site, free of cost
 - The firm should be able to give demonstration of the item if required within 5 days of requisition.
 - The firm must supply the item within 20 days after confirmation of order.
 - **The successful bidder has to submit performances bank guarantee of 10% of the order value valid upto warranty period.**
 - The payment will be made electronically viz RTGS/NEFT against delivery and satisfactory installation of the item. The following information should be also mentioned in the quotation.
 - Name of the beneficiary:
 - Account No. of the beneficiary:
 - IFCS code of the bank/branch:
 - The quantity can be increased or decreased to any extent depending upon the actual requirement.
 - The Professor & Head of the Department reserves the right to cancel/reject full or any part of the rate enquiry which generally do not fulfil the conditions stipulated in the rate enquiry.
 - Payment will be made after receiving the items in good condition and satisfactory installation.
 - No extra transportation charge will be given to the firm.
 - The firm must supply the items within the date mentioned on the supply order copy.
- Thanking you


(Dr. Renu Dhingra)

Professor

Computer facility: With request for uploading the same in AIIMS websites for the period of 15 Days.

Specification for refrigerators

Refrigerator should have:- (Minimum capacity 300-350 L Numbers :4)

1. Should be of vertical type
2. Minimum capacity 300-350 L
3. Double door
4. Digital inverter
5. Should have a built in condenser
6. Should have puff insulation
7. Should use CFC free refrigerant
8. Should provide optimal automatic Voltage Stabilizer with Automatic Low High Cut-Off & TDR 13.
9. Should be energy efficient
10. Power: 220-240V 16A/3680W
11. *Warranty and guaranty as per AIIMS rule*
12. *Vendor should have service centre at Delhi and NCR*

Refrigerator should have:- (Minimum capacity 400- 450L numbers :2)

1. Should be of vertical type
2. Minimum capacity 400- 450L
3. Double door
4. **Should maintain temperature of $\leq -20^{\circ}\text{C}$ in the freezer compartment ($\pm 5^{\circ}\text{C}$) and $= 0^{\circ}\text{C}$ ($\pm 4^{\circ}\text{C}$) in the refrigerator compartment**
5. Should have a built in condenser and Digital inverter
6. Should have puff insulation
7. Should use CFC free refrigerant
8. Should provide optimal automatic Voltage Stabilizer with Automatic Low High Cut-Off & TDR 13.
9. Should be energy efficient
10. Power: 220-240V 16A/3680W
11. *Warranty and guaranty as per AIIMS rule*
12. *Vendor should have service centre at Delhi and NCR*

Renu Dhirga
Dr. Renu Dhirga
Professor
Department of Anatomy
A.I.I.M.S., New Delhi-29

Saravjee Ray

hmt Kar

Abhang