

**DR. RAJENDRA PRASAD CENTRE FOR OPHTHALMIC SCIENCES**  
**All-India Institute of Medical Sciences**  
**Ansari Nagar, New Delhi-29**

Ref. No.SO/RPC/Proprietary/Slit Lamp/2014-15

Dated: 02.03.2014

**Subject: Purchase of equipment Slit Lamp – 06 Nos. for Dr. R.P.Centre at AIIMS, New Delhi-29 on proprietary basis- Inviting comments thereon.**

\*\*\*\*\*

The request received from respective faculties of Dr. R.P.Centre AIIMS for the purchase of subject cited equipment from M/s. Haag Streit International, USA on proprietary basis. The proposal submitted by M/s. Haag Streit Internation, USA and PAC certifications are attached & uploaded on website.

The above documents are being uploaded for open information to submit objections, comments, if any, from any manufacturer regarding proprietary nature of the equipment/item within 15 days from the date of issue/uploading of the notification giving reference **SO/RPC/Proprietary/Slit Lamp/2014-15**. The comments should be sent to Stores Officer, Dr. R.P.Centre at AIIMS on or before **17.03.2015 upto 04.30 P.M.**, failing which it will be presumed that any other vendor is having no comment to offer and case will be decided on merits.

**Yours faithfully,**

**STORES OFFICER (RPC)**

**Encl: Related documents enclosed.**

- 1. PAC Certificate enclosed.**
- 2. Specification of equipment.**

**Technical Specification:**

1. Should have illumination from top using tungsten filament bulb capable of giving illumination intensity up to 600,00Lux.
2. Magnification from 6.3X to 40X having continuous zoom
3. Slit length 0.2 to 8mm
4. Slit length 0.2 to 8mm
5. Eyepieces of 12.5X
6. Facility to tilt the slit image up to 20 deg. Should be available
7. Should have applanation tonometer
8. Should have the facility for stereoscopic examination of fundus where in Angle on stereoscopic observation can be reduced from 13 Deg. To 4.5 Deg. To have better view of eyes with small pupil or high myopia.
9. Facility for viewing into the microscope at 20 deg to horizontal to enable the examiner to keep his head in fatigue free position.

Slit lamp should be mounted on spring balanced instrument stand

Date: February 27, 2015

To,  
The Stores Officer,  
Dr, RP Centre, AIIMS,  
Ansari Nagar,  
New Delhi-110029.

Dear Sir,

**Sub: Proprietary Certificate**

This is to certify that Slit Lamp BQ 900 is a proprietary product of Haag Streit, Switzerland. The Slit Lamp BQ 900 has a number of features of that are proprietary and unique, that are wholly property of Haag Streit AG, Switzerland. Surgeons and the patients have benefited from the accuracy, reliability and customization, the market leading technology.

Below is the list of the unique features which include but not limited to:

**The Microscope**

The convergent Stereo view of the BQ 900 offers a very large binocular field of view with high resolution and depth of focus. Eyepieces adjustable to all pupillary distances allow for a fatigue-free and relaxed examination, even during long examinations.

**Stereo Variator**

The Slit Lamp also has been identified by its unique option called Stereo Variator (Optional) which reduces the 13° angle to 4.5°. Thus an enlarged binocular field of view is possible even with myopia or through small pupils. This also aids the observation, without additional lenses, of the peripheral fundus and vitreous. It improves fundus viewing for imaging applications.

**Illumination**

The well established tungsten slit illumination upto 6,00,000 Lux brightness paired with precision mechanics and high quality optics establish an excellent slit image with highest brightness from 8 mm length and width to a hairline. The possibility to rotate the slit 90° around the optical axis on either side, tilting it up to 20° and decentring it, lets you find the right adjustment for all examination situations.

**Inclined Eyepiece Adaptor**

The Slit Lamp is designed with simple but effective device which allows the eyepieces to be inclined upwards by 20° from their traditional horizontal position, thereby reducing neck strain in allowing a more natural viewing position to the user.

Yours faithfully,



Marco Obbens  
International Sales Director



Dr. Reiner Herrmann  
International Marketing Director