ALL INDIA INSTITUTE OF MEDICAL SCIENCES DEPARTMENT OF ANATOMY

Ansari Nagar, New Delhi-110029

RE/ 02 /Anat/Store/15-16

Dated :-25 / _6 / 2015

Dear M/s,

A short rate enquiry is hereby floated/invited for the following item required by Department of Anatomy, AIIMS, New Delhi-110029.

S.No	Name of Items
1	RNA work station
2	Mini centrifuge
3	Metabolic cages

Firm submit the following:

- 1. Firm should submit price Reasonable Certificate
- 2. Firm should submit the under taking on the letter that the firm has not been black listed
- 3. Firm should submit the under taking on the letter that the firm has no vigilance case pending against Supply /firm

Terms & Conditions:

- The quotations should be submitted to the Professor & Head, Department of Anatomy, Room No. 1010, AIIMS, New Delhi-29 and sent by post/per bearer by 20th July, 2015 up to 12.00 P.M. Quotation should be sealed in an envelope and the "name of the item" clearly written on top of the envelop.
- The make of the article offered should invariably be quoted. Quotation should be typed written in ink. All overwriting and erased entries will be deleted from the quotation. The rates should be valid for at least THREE MONTHS.
- VAT/Sales tax or any other kind of tax(s) must be separately mentioned against each item. In case no sales tax is chargeable, prices must be quoted as NET.
- Warranty/Guarantee: 1 year comprehensive warranty.
- Installation: The firm install the equipment on the site, free of cost
- The firm should be able to give demonstration of the item if required within 5 days of requisition.
- The firm must supply the item within 20 days after confirmation of order.
- The successful bidder has to submit performances bank guarantee of 10% of the order value valid upto warranty period.
- The payment will be made electronically viz RTGS/NEFT against delivery and satisfactory installation of the item. The following information should be also mentioned in the quotation.
- Name of the beneficiary:
- Account No. of the beneficiary:
- IFCS code of the bank/branch:
- The quantity can be increased or decreased to any extent depending upon the actual requirement.
- The Professor & Head of the Department reserves the right to cancel/reject full or any part of the rate enquiry which generally do not fulfil the conditions stipulated in the rate enquiry.
- Payment will be made after receiving the items in good condition and satisfactory installation.
- No extra transportation charge will be given to the firm.
- The firm must supply the items within the date mentioned on the supply order copy.

Thanking you

(Dr.Renu Dhingra) Professor Renu Dhingra

Computer Recipity: With request for uploading the same in AIIMS websites for the period of 10

Days ... New Delhi ?

Compact Micro spin Centrifuge should have:

- 1. Speed range 1,4000-15000 rpm and option for 12 places and 16 place rotor to spin micro centrifuge tubes or PCR tubes
- 2. G value should be more than 14,000
- 3. Automatic RPM/RCF Conversion.
- 4. Rotor bowl should be made of Metal / Stainless steel.
- 5. Rotor and rotor lid should be chemical resistance.
- 6. Separate short spin button for quick mixing.
- 7. Timer 15 sec to 99 mins and continuous mode will be prefer
- 8. Option to spin different size tubes like 0.2, 0.4, 0.5 ml & 0.6 ml tubes. And other accessories should be quoted as optional.
- 9. Compact centrifuge should have Compact Centrifuge should have Acceleration and braking time not more than 13-15 sec.
- 10. Automatic door opening after the run
- 11. Noise level not more than 52dBA..
- 12. All rotors autoclavable.
- 13. Confirms to European CE/ IEC/ 1010-2-020 safety regulations.
- 14. Compact Centrifuge should have IVD certificate
- 15. Auto Opening of lid after the run.
- 16. Quoted model should have good no. of installation. And End user certificates should be enclosed along with Bid.
- 17. Guaranty/ warranty, AMC and CMC according to AIIMS rules

Renu Dhingra
Dr. Renu D

RNA work station specifications

The RNA work station should have

- 1. white fluroscent light and 254nm UV lamp with 15 min. timer provides for decontamination
- 2. Compact footprint—less than 5 sq. ft. turns any area of the lab into a PCR clean room
- 3. dimensions not more than 2ft x 2ft x 3.0ft
- 4. seprace switch for all lights.
- 5. prefilter included in the work station.
- 6. rust free property
- 7. aluminium surface
- 8. Warranty and gurranty accoding to AIIMS rules

Remu Dhingra

Dr. Renu Dhingra

Dr. Renu Dhingra

Department of Anatomy

Department of Anatomy