



**DEPARTMENT OF MICROBIOLOGY**  
**ALL INDIA INSTITUTE OF MEDICAL SCIENCES**

S. No: 15/Micro/2014-15

Dated: 05.02.2015

*A short rate enquiry (LIMITED TENDER) is floated/invited for the purchase of below mentioned item as per the specifications.*

Name of item	Qty.
<b>BOD INCUBATOR</b>	<b>01</b>

*Specifications:*

1. **Temperature Range** - 0°C to 60°C with an accuracy of  $\pm 0.5^\circ\text{C}$
2. **Temperature Controller** - Program P.I.D. Controller 5 pattern 18 step
3. **Temperature Regulator** - Triac Controller
4. **Timer Range** - 99 min 59 sec with Digital Display
5. **Interior Construction** - Stainless Steel Plate
6. **Exterior Construction** - Steel plate with powder coating
7. **Insulation** - Glass Wool
8. **Doors** – Double door magnetic gum packing
9. **Inner Door**- High temperature safety glass of thickness 5 mm
10. **Shelves** - Two with adjustable type Stainless Steel shelves
11. **Safety Device** - Over Temperature limit / Over Current Glass Fuse
12. **Freezer Compressor** - Air Cooling unit system
13. **Capacity** - 81 Litres
14. **Inner Size** - 450 x 400 x 450 mm (W x D x H)
15. **Outer Size** - 625 x 740 x 950mm (W x D x H)
16. **Heater capacity** - 650W
17. **Warranty** - Minimum of 2 years.

## **Terms & Conditions:-**

1. Firm should submit VAT registration certificate.
2. Authority letter from manufacturer/principal should be annexed with quote.
3. The firm is requested to submit the undertaking on the firm's letterhead that the firm has **not been black listed** in the past by any hospital/organization.
4. The firm is requested to submit the undertaking on the firm's letterhead that the firm has **no vigilance case/CBI case pending against him/suppliers.**
5. The firm is requested to submit **Fall clause undertaking** on the firm's letterhead.
6. Firm should clearly mention the **brand name** for the product.
7. Literature of the original catalogue of the product.
8. The sealed quotation should reach in **Room # 2080**, Microbiology department under two bid system on or before **12/02/2015 upto 12:00 noon** after entry being made in register.
9. The Technical Bids may be opened on **12/02/2015 at 12:30 pm** in the undersigned office.

Tender should be submitted in **TWO BID SYSTEM** containing two parts as detailed below:-

**Part I : Technical Bid in one sealed envelope**

**Part II : Price Bid in one sealed envelope**

The sealed envelopes should then be put in one outer envelope indicating thereon:

- a) *Reference No. of the short rate enquiry*
- b) *Item Name*
- c) *Name of the Firm*

*Those firms who fulfil/submit pre-qualification documents and meet the technical specifications shall be called for demonstration of the equipment, if felt necessary. The final technical selection will be made only after satisfactory demonstration of quoted item/model.*

**DR. IMMACULATA XESS**  
Professor