

**JAI PRAKASH NARAIN APEX TRAUMA CENTER
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
RAJ NAGAR, NEW DELHI : 110 029**

TC-983/Surg/14-15/FSC

Dated : 14.01.2015

**SUBJECT : Double combination OT Ceiling Light with LED Technology
and Centrally Mounted HD camera system, Medical Grade
Monitor.**

AMENDMENT MADE IN THE SPECIFICATION

**The following changes has been made in the specification which has
already been published.**

- 1. Two Dome Lights instead of Three Domes as the Ceiling height is
low.**
- 2. Monitor should be on a separate arm independent of OT light to
be installed on the Ceiling.**

**The above amendment is issued without prejudice to other terms &
conditions.**

**(K.D. SHARMA)
SR. STORE OFFICER**

**JAI PRAKASH NARAIN APEX TRAUMA CENTER
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
RAJ NAGAR, NEW DELHI : 110 029**

CORRIGENDUM

Please refer to our following TENDER published in various newspaper and through our website at www.aiims.edu / www.aiims.ac.in and www.tender.gov.in. and the undersigned is directed to notify to all concerned that as per suggestions of various bidders , the specifications of the following tenders have been amended and published through our above website. The list of amendment can be downloaded from the above website or can collect from the Stores Section (Basement), JPNA Trauma Centre, A.I.I.M.S., Raj Nagar New Delhi-29 upto 03.02.2015 upto 4.00 PM.

1. TC-983/Surg/14-15/FSC (OT Ceiling Light with LED Technology and Centrally mounted HD Camera System, Medical Grade Monitor)

The last date of submission of tender documents is also extended upto 04.02.2015 at 12.30 P.M.

The Technical bids will be opened on 05.02.2015 at 10.30 A.m.

The above corrigendum is issued without prejudice to other terms & conditions.

**(SR. STORES OFFICER)
JPNATC**

The copy of this corrigendum may be forwarded to Computer Facility, Main AIIMS for uploading in the AIIMS website which has already been forwarded through their e-mail ID i.e. contentprovideraiims@gmail.com.