

COMPREHENSIVE RURAL HEALTH SERVICES PROJECT CENTRE FOR COMMUNITY MEDICINE ALL INDIA INSTITUTE OF MEDICAL SCIENCES BALLABGARH – 121004 (HR) / Ph. 0129-2241362/Fax. 0129-2211227

S.R.E.No.26/15-16/CRHSP-AIIMS /Ballabgarh

Dated the: 26th November, 2015

For Purchase of : (i) (Incubator for incubating the Elisa plate):- Qty.-Approx. 01 Nos.

The sealed quotations are invited by post/per bearer (under two bid system) in two separate sealed covers duly marked as "PRICE BID & TECHNICAL BID" (as the case may be) and both packed in other single envelope mentioning name of the quoted item for the supply of item detailed below which are required for CRHS Project, Ballabgarh.

The quotations duly signed and marked as "SRE No.26/CRHSP/Ballabgarh/15-16" (PRICE BID & TECHNICAL BID separately) should reach the office of the Professor In-charge, CRHSP, Ballabgarh on or before 11/12/2015 at 3.00 P.M. All quotations should be type written or written with indelible ink. All over writing and erased entries will not be considered and treated as deleted entries from the quotations. The rates should be valid at least for one year. A clause by clause compliance statement should be submitted, failing which the quotation will not be considered.

The tenders are required to enclose with the quotation an Earnest Money of Rs.5,000/- (Rupees Five Thousand only) through a demand draft/pay order drawn in favor of the "Director, AllMS", failing which the quotation will not be considered. In case of failure to comply with the supply order in terms of the contract, purchase will be made at the cost and risk of the contractors and or liquidated damages may be recovered from them. The period within which the goods can be supplied from the date of issue of the supply order should be specified.

The offers submitted as quality A, B, C, & D etc., will be summarily rejected, any alteration/over writing in the specification / rates shall not be considered at all.

The opening of quotations/tenders (technical bid only) will be intimated to the bidders separately. The quantity shown in the Short Rate Enquiry can be increased or decreased to any extent depending upon the actual requirement.

Technical Specification for incubator for incubating the Elisa Plate:-

- 01. The incubator should be suitable for incubation of ELISA plates of TTI markers
- 02. Dimensions should be approximately 355 × 355 × 355 mm
- Should be sturdy, double walled construction with complete inner chamber made of Anodized Aluminum or Highly Polished Stainless Steel.
- 04. Gap between the walls should be filled with special grade insulation material for proper insulation and to avoid heat losses.
- 05. Should have inner chamber fabricated with ribs for adjusting shelves to convenient height.
- 06. Should be supplied with 2 or 3 removable shelves.
- 07. Heating element should be placed at the bottom and side ribs for uniform temperature all over the space.
- 08. Should have the temperature control by thermostat from 20°C above ambient to 50°C ±0.5°C. Temperature control knob should be graduated in degrees centigrade.
- 09. Should be provided with Air ventilators port on sides at top for ventilate fumes & to assist convection process.
- The equipment should be provided with a panel having a thermostat control knob, ON/OFF switch and indicator light.
- 11. Should be supplied with cord and plug. Suitable to operate on 220 V single phase, 50 Hz, AC supply.

Terms and Conditions:-

Comprehensive warranty for Two year from the date of purchase. Standard AllMS terms and conditions apply.

PROFESSOR-IN-CHARGE, CRHS PROJECT, BALLAGARH

The firm should submit the following:-

- 1. A clause by clause compliance statement should be submitted.
- 2. Price Reasonable Certificate.
- 3.An undertaking on the official letterhead of the company that the firm has not been black listed.
- 4.An undertaking on the official letter head of the company that no Vigilance case is pending against the firm.

#