



COMPREHENSIVE RURAL HEALTH SERVICES PROJECT
CENTRE FOR COMMUNITY MEDICINE
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
BALLABGARH - 121004 (HR) / Ph. 0129-2241362/Fax. 0129-22112 27

S.R.E.No.12/16-17/CRHSP-AIIMS /Ballabgarh

Dated the: 06.12.2016

For Purchase of : Fetal Doppler Machine - Qty. 40 Nos.

The sealed quotations are invited by post/per bearer (under two bid system) in **two separate sealed covers duly marked as "PRICE BID & TECHNICAL BID"** (as the case may be) and both packed in other single envelope mentioning name of the quoted item for the supply of item detailed below which are required for CRHS Project, Ballabgarh.

The quotations duly signed and marked as **"SRE No.12/CRHSP/Ballabgarh/16-17" (PRICE BID & TECHNICAL BID separately)** should reach the office of the Professor In-charge, CRHSP, Ballabgarh on or before **22nd December, 2016 at 12.00 Noon**. All quotations should be type written or written with indelible ink. All over writing and erased entries will not be considered and treated as deleted entries from the quotations. The rates should be valid at least for one year. **A clause by clause compliance statement should be submitted, failing which the quotation will not be considered.**

The tenders are required to enclose with the quotation an **Earnest Money of Rs.5,000/- (Rupees Five Thousand only)** through a demand draft/pay order drawn in favor of the **"Director, AIIMS"**, failing which the quotation will not be considered. In case of failure to comply with the supply order in terms of the contract, purchase will be made at the cost and risk of the contractors and or liquidated damages may be recovered from them. The period within which the goods can be supplied from the date of issue of the supply order should be specified.

The offers submitted as quality A, B, C, & D etc., will be summarily rejected, any alteration/over writing in the specification / rates shall not be considered at all.


The opening of quotations/tenders (technical bid only) will be intimated to the bidders separately. The quantity shown in the Short Rate Enquiry can be increased or decreased to any extent depending upon the actual requirement.

Terms and Conditions:-

Standard AIIMS terms and conditions apply.

The firm should submit the following:-

1. **A clause by clause compliance statement should be submitted.**
2. Price Reasonable Certificate.
3. An undertaking on the official letterhead of the company that the firm has not been black listed.
4. An undertaking on the official letter head of the company that no Vigilance case is pending against the firm.


(DR. SHASHI KANT)
PROFESSOR-IN-CHARGE,
CRHS PROJECT, BALLABGARH
आचार्य एचएम विभागाध्यक्ष/Professor & Head
सामुदायिक चिकित्सा विज्ञान केंद्र
Centre for Community Medicine
अ. मा. आ. सं. नई दिल्ली/AIIMS, New Delhi



COMPREHENSIVE RURAL HEALTH SERVICES PROJECT
CENTER FOR COMMUNITY MEDICINE
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
BALLABGARH – 121004 (HR)

Ph. 0129-2241362
Fax. 0129-2211227

Technical Specification for Fetal Doppler

- *Should be Compact, Pocket Type and Light Weight .*
- *Large LCD Display for Display of Fetal Heart Rate*
- *Back Light for Better Visibility*
- *Display Signal Quality, Low Battery et.,*
- *High Sensitivity Doppler probe of frequency 2.0 MHz*
- *Ultra Sound intensity <10mw/cm².*
- *Auto Shut Off Facility to save Battery Power*
- *Built-in Speaker with Output not less than 1.0W*
- *Should work on Rechargeable Batteries with Minimum Battery Time of 300 minutes.*
- *Battery Charger and Spare set of Batteries AA type to be supplied.*
- *Volume control facility and Audio Output for Ear Phone should be available*
- *Heart Rate Range should be from 50 to 240 bpm with accuracy of +/-2%*
- *Should be Eater Proof Body*
- *Should have facility FHR Data transfer to PC.*
- *Doppler probe should be Light Weight with Holder Facility when not in use.*
- *Should be supplied with Carrying Case.*
- *Should have FDA/CE Certificate for International Quality Standard.*