

DEPARTMENT OF HEMATOLOGY
A.I.I.M.S. NEW DELHI

18.11.2015

CORRIGENDUM

Dr. P. Kotwal
Chairman, Computer Facility, AIIMS

Sub: For giving in website Rate Enquiry No. 4/Haem/2015-16 for purchase of -20°C Deep freezer.

Dear Sir,

Enclosed the specification for -20°C Deep freezer. Invite the Sealed quotations for -20°C Deep freezer as per specifications given below. Quotations to be send to Dr. Renu Saxena, Prof. & Head, Department of Hematology A.I.I.M.S, New Delhi -110029. The last date for receiving the sealed quotation is 4th December 2015. Kindly include in AIIMS Website.

Thanking you,

Yours Sincerely

DR. RENU SAXENA
PROF. & HEAD DEPT. HEMATOLOGY

DEPARTMENT OF HEMATOLOGY, A.I.I.M.S.

SPECIFICATION FOR -20°C Deep Freezer

- Should have Vertical Upright model
- Should have Microprocessor touch screen control with digital display
- Should have Temperature performance -20°C
- It should have external double door
- Should have 500 to 550 litres
- Inner chamber should be made of high quality stainless steel
- Control panel at the top of cabinet with temperature warning light
- Should have CFC free refrigerant
- Battery back-up for alarm in case of power failure
- Audio/visual alarm and adjustable high and low alarm
- Should have Air cooled condenser
- Should have Static cooling system
- Should have Hermetically sealed compressor
- PT 100 sensor mounted inside the chamber
- It should have minimum 3 to 4 adjustable shelves
- Should have ISO, CE/FDA certified
- Should have voltage stabilizer and portable data logger with freezer
- Should have minimum one year warranty.
- **It should be two bid system Technical bid & Price bid Should be put in separate sealed envelope.**