



COMPREHENSIVE RURAL HEALTH SERVICES PROJECT
CENTRE FOR COMMUNITY MEDICINE
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
BALLABGARH - 121004 (HR)

Ph. 91- 0129-2241362
Fax. 91-0129-2211227

No.F.R.E.03/CRHSP-AIIMS/Ballabgarh/2016-17

Dated the 19.8.2016

To,

Subject: Quotation for purchase of Infusion Syringe Pump (Quantity 5 Nos.) at CRHSP, Ballabgarh Hospital.

Sir,

You are requested to please give your quotation for purchase of Infusion Syringe Pump at CRHSP, Ballabgarh Hospital. Quotation should reach the office of the undersigned latest by 05.9.2016. Specification is as under:-

Normal Flow rate range:		Syringe Capacity	Flow Rate
		10ml	0.1 -300 ml/hr
		20ml	0.1 -500 ml/hr
		30ml.	0.1 -700 ml/hr
		50/60 ml	0.1 -900.9 ml/hr
Accuracy	Mechanical Accuracy		+1%
	Overall Accuracy*		+2% (measured time>1hr, min 0.2ml. excluding syringe tolerances.
Occlusion Pressure Limits:	Low		0.4+/-0.08kg/cm2 (bar)
	Medium		0.8+/-0.20KG/cm2 (bar)
	High		1.2+/-0.25 Kg/cm2 (bar)
Operating conditions:	Ambient Temperature		10 to 40o C
	Relative Humidity		20 to 90% (on condensation)
	Atmospheric Pressure		70 to 106 kpa
Power:	Power Source		190-280 VAC, 50/60Hz
	Power Consumption		12 VA Maximum
	Fuse		2A

Terms and conditions :

Comprehensive warranty for two years from the date of purchase.
Standard AIIMS terms and conditions apply.

The responsibility for warranty will be entirely with the local dealer for the full period

Thanking you,

for

(Dr. Shashi Kant)
Professor-in-charge,
CRHS Project, Ballabgarh

Dr. Rakesh Kumar
Assistant Professor
C.R.H.S. Project, Ballabgarh
Centre for Community Medicine
All India Institute of Medical Sciences
New Delhi