

DEPARTMENT OF MICROBIOLOGY
ALL INDIA INSTITUTE OF MEDICAL SCIENCES

Rate Enquiry No. 3 /Micro/2016-2017

Dated 29 -10-2016


Quotations are invited for the supply of items mentioned below which are required for Microbiology Department AIIMS.


1. Shaker Incubator – 1 No.
2. Binocular Compound Microscope – 2 Nos.
3. Water Bath – 1 No.
4. Weighing Balance – 1 No.
5. Horizontal Gel Electrophoresis unit – 1 No.
6. Electrophoresis Power Pack – 1 No.
7. pH Meter – 1 No.
8. LED Fluorescent Microscope – 1 No.
9. Electronic Balance – 1 No.
10. Autoclave – 1 No.

The quotations should be submitted to Prof. & Head Department of Microbiology Department Room No.2060 IInd Floor, Teaching Block, AIIMS, New Delhi by 19-11-2016 upto 4.00 pm. The quotations will be opened on 21-11-2016 at 3.00 pm Room No.2093. All quotations should be typewriter or in ink. All over writing and erased entries will be deleted from the quotations. The quotation should be duly sealed in an envelope marked R.E.No. 3 /Micro/2016-2017.

Terms & Condition:

1. The quotation for each item should be submitted in a separate with the name of the item clearly written on top of the envelope.
2. The Rate Enquiry must quote for Five Years Comprehensive Warranty.
3. Authority letter from manufacturers/principal for supplying item.
4. Price reasonable certificate.
5. An undertaking that the firm has been blacklisted in the past by any Institution, Government/Private.
6. The names & contact numbers of AIIMS/Govt. Departments where the supplier/dealer has supplied the equipment (user list).
7. Information for electronic payment viz RTGS/NEFT – Name of the beneficiary, Account number of the beneficiary, IFSC code of the bank/branch.
8. If applicable, VAT/sales tax or any other kind of tax must be mentioned separately against each item.
9. The firm must be able to install the equipment on the site, free of cost.
10. The firm must supply the items within the date mentioned on the supply order copy.


Dr. Arti Kapil
Professor of Microbiology
Officer Incharge Store


Dr. Gita Satpathy
Prof. & Head
Department of Microbiology

WATER-BATH Specifications:

1. Should be provided with Tank : High grade SS tank
2. Should be provided with Capacity : 5 to 15 litres
3. Should be provided with Temperature range : Ambient +5°C to 99.9°C
4. Should be provided with Temperature accuracy : $\pm 0.2^\circ \text{C}$
5. Should be provided with Digital display and selection of temperature
6. Should be provided with Timer : 0.1 Hour to 99.9 Hours
7. Should be provided with continuous Safety protection : User settable and fixed thermal cut-out
8. Should be provided with Display : Digital display of temperature
9. Should be provided with Alarm : Low level water alarm, safety shut down audible/visual alarm
10. Should be provided with Operating voltage : $230 \pm 10 \text{VAC}$, 50 Hz.
11. Should be provided with Tank design for easy cleaning and draining
12. Should be provided with weight 5-15 kg
13. Should be provided with comprehensive warranty of 5 year
14. Should be ISO 9001 and 14001 certified. The instrument should be an original make from the company and not assembled locally

Certified that the specifications are broad based, general in respect to the requirement and not suit to any particular firm/brand.

Checked by

Ramchandra

6/2/2

Dr. TANU SAGAR, MD

Department of Microbiology
All India Institute of Medical Sciences
Ansar Nagar, New Delhi-29

all right

Dr. TANU SAGAR, MD
सहायक आचार्य/Assistant Professor
सूक्ष्म जीव विज्ञान विभाग/Deptt. of Microbiology
अखिल भारतीय आयुर्विज्ञान संस्थान/A.I.I.M.S.
अंसारी नगर, नई दिल्ली/Ansan Nagar, New Delhi-29