

DEPARTMENT OF MICROBIOLOGY  
ALL INDIA INSTITUTE OF MEDICAL SCIENCES

Rate Enquiry No. 3 /Micro/2016-2017

Dated 29 -10-2016


Quotations are invited for the supply of items mentioned below which are required for Microbiology Department AIIMS.


1. Shaker Incubator – 1 No.
2. Binocular Compound Microscope – 2 Nos.
3. Water Bath – 1 No.
4. Weighing Balance – 1 No.
5. Horizontal Gel Electrophoresis unit – 1 No.
6. Electrophoresis Power Pack – 1 No.
7. pH Meter – 1 No.
8. LED Fluorescent Microscope – 1 No.
9. Electronic Balance – 1 No.
10. Autoclave – 1 No.

The quotations should be submitted to Prof. & Head Department of Microbiology Department Room No.2060 IInd Floor, Teaching Block, AIIMS, New Delhi by 19-11-2016 upto 4.00 pm. The quotations will be opened on 21-11-2016 at 3.00 pm Room No.2093. All quotations should be typewriter or in ink. All over writing and erased entries will be deleted from the quotations. The quotation should be duly sealed in an envelope marked R.E.No. 3 /Micro/2016-2017.

Terms & Condition:

1. The quotation for each item should be submitted in a separate with the name of the item clearly written on top of the envelope.
2. The Rate Enquiry must quote for Five Years Comprehensive Warranty.
3. Authority letter from manufacturers/principal for supplying item.
4. Price reasonable certificate.
5. An undertaking that the firm has been blacklisted in the past by any Institution, Government/Private.
6. The names & contact numbers of AIIMS/Govt. Departments where the supplier/dealer has supplied the equipment (user list).
7. Information for electronic payment viz RTGS/NEFT – Name of the beneficiary, Account number of the beneficiary, IFSC code of the bank/branch.
8. If applicable, VAT/sales tax or any other kind of tax must be mentioned separately against each item.
9. The firm must be able to install the equipment on the site, free of cost.
10. The firm must supply the items within the date mentioned on the supply order copy.


  
Dr. Arti Kapil  
Professor of Microbiology  
Officer Incharge Store


  
Dr. Gita Satpathy  
Prof. & Head  
Department of Microbiology

Specification for Electronic weighing Balance

1. Weighing range should be at least 2 mg to 200 g
2. Should have pan diameter of 70-90 mm
3. Dimensions should not be more than 190 x 260 x 310 (w x d x h) mm.
4. The readability range should be 0.0001 and linearity of 0.2 mg
5. The mass should be < 5 kg.
6. Should be provided with a compatible adapter, and power requirements should meet Indian standards
7. Should be provided with 5 years comprehensive warranty
8. Should feature easy operation, fast readouts, precision weighing and reliability.
9. User list must be submitted along with offer.
10. Company should be ISO/ISI/EU(European Union) certification.

Certified that the specifications are broad based, general in respect to the requirement and do not suit any particular firm/brand.

  
Prepared by

  
Checked by

