

Department of N.M.R.
All India Institute of Medical Sciences
Ansari Nagar, New Delhi – 110 029 INDIA

Short Rate Enquiry No. NMR/Dept/Budget/2019-20/

Dated: 05.03.2020

Subject:- Short Rate Enquiry for purchasing **Digital weighing analytical balance**

A short rate enquiry is hereby floated/ invited for purchasing “Digital weighing analytical balance” required by Department of N.M.R.

Date/Time of Closure 19.03.2020; 3.00pm

Technical Specification for Digital weighing analytical balance

1. Weighing capacity range: 0.01 (lower range/ minimum weight) to 60gm or above (upper range)
2. The balance should have Automatic Internal Adjustment
3. The quoted system should have Touch Screen Panel with graphic display in the balance
4. The quoted system should have Water and acid proof cabinet
5. The quoted system should have Easy Calibration adjustment without tools
6. Density determination kit for the balance with weighing pans
7. Pipettes MOC should have acid proof, corrosion resistant and autoclavable (n=3) with adjustable volume
8. [Volume upto (i) 10 μ l (ii) 100 μ l and (iii) 1000 μ l] for liquid sample
9. Warranty 05 years from the date of satisfactory installation (02 years comprehensive warranty + 03 year CAMC Warranty)