

All-India Institute of Medical Sciences
Ansari Nagar, New Delhi-29
(RESEARCH SECTION)

Ref. No. 23/Prop/Anat./SCY/21-22/RS

Dated:- 14.3.2022

Subject: Procurement of Ultramicrotome – 01 No. on proprietary basis - Inviting comments thereon.

The request has been received from Dr. Subhash Chandra Yadav, Additional Professor, Dept. of Anatomy for procurement of the subject item from M/s Leica Microsystems (A Division of DHR Holding India Pvt. Ltd.) . (Mfg: M/s. Leica Mikrosysteme Vertrieb GmbH.) on proprietary basis. The proposals submitted by M/s Leica Microsystems (A Division of DHR Holding India Pvt. Ltd.) . (Mfg: M/s. Leica Mikrosysteme Vertrieb GmbH.) and Performa Invoice and Departmental PAC certifications are attached.

The above documents are being uploaded for open information to submit objections, comments, if any, from any manufacturer regarding proprietary nature of the equipment/item within issue of 15 days giving reference 23/Prop/Anat./SCY/21-22/RS . The comments should be received by Office of Stores Officer (RS), Research Section at AIIMS on or before 30.3.2022 up to 12:00 P.M., failing which it will be presumed that any other vendor is having no comment to offer and case will be decided on merits.


ASST. STORES OFFICER (RS)

Encl: Related documents enclosed.

1. PAC Certificate enclosed.
2. Performa Invoice

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI-110029

2

RESEARCH SECTION

PROPRIETARY/SPECIFIC BRAND GOODS CERTIFICATE

1. Item/Type/Model No. required alongwith specification. : Ultramicrotome / sample preparation
Model No - Leica EM UC-7
2. Is the item a spare parts attachment or accessory for an existing equipment. : NO
3. Name of the manufacturers/supplier of the item proposed by the Indentor. : Leica Microsystems Vertrieb GmbH / DHR Holding India Pvt. Ltd.
4. Are they sole manufacturers/sole distributors of the item. : Yes
5. Is there any other item with similar/ equivalent specification available in the market to meet the job requirement envisaged. If the answer is yes, why the same can't be procured. Demanding officer should bring out comparative functional advantages/cost effectiveness of the recommended item from these offered by other. : NO
6. What were the efforts made to locate alternative source of supply or use other substitutes. : NA
7. Why open/limited tender can't be resorted to, for locating alternative source. : NA
8. Are the proprietary items certifying that the rates are reasonable or not. : Yes
9. Any other justification for procuring item from single source. : They are the sole distributor for this equipment.

Signature of Indentor
(Demanding Officer)

COUNTERSIGN
(Head of the Department)

I certify that the item at Sr. No. 1 above is required to be procured on single tender basis as the source of supply is definitely known/the specified brand proposed was advantages in meeting our functional requirements and limited tender system could be dispensed with as they would serve no useful purpose in this particular case.

(Strike out whichever is not applicable)

Subhas Mehta, MS
Director Professor & Head (Anatomy)
VMMC & Safdarjung Hospital
New Delhi - 110029

Subhas

Subhas

AK Mehta

Pratima

Pratima

Pratima

(Electron Microscope Facility, Department of Anatomy, AIIMS New Delhi)

Item No. 5; Qty required - 01 ; Approximate Cost – Rs.40.00 lakhs (Research Section)

Technical Specification for Ultramicrotome

1. Should have cutting transmission with vibration decoupled gravity stroke & built in Anti-vibration system
2. Illumination should have independently controllable
3. Should have built in intensity control for Ioniser with warm hand rests
4. Should have stereomicroscope with 10x to 75x magnification & integrated ergo wedge
5. Should have eucentric microscope movement with defined click positions for glass knife and diamond knife
6. Basic Instrument should have 180° swivelling microscope carrier.
7. Should have 4 x LED for top light, spot light, back light and specimen trans illumination
8. Should have fully motorized knife stage E-W movement & N-S step movement
9. Should have 10.4" touch screen controlling the ultramicrotome
10. Should have display of countdown, E-W measurement & knife usage monitoring
11. Should have standard sectioning counter/feed totalizer and countdown mode for exact trimming
12. Should have E-W measurement with auto trim & rocking mode trim function
13. Should have user, sectioning, knife parameters and grid box parameters download via USB (logbook)
14. Should be supplied with two Diamond knives ultra 45° with at least 2.0mm width, two diamond cryo knives, and two complete active anti-vibrational table from the manufacturer or recognised by principle
15. Should quote required spares & consumables along with the quotation
16. All the above-mentioned modules and required parts must be covered during the period of Warranty and CAMC.
17. The vendor should provide a rate-list of all relevant consumables and spares and the price should be frozen for the period of warranty. Any part required for the functioning of these equipment that has not been included in this rate list must be provided by the vendor free of cost.
18. The system should be certified from European CE / UL / ISO 9001/Bureau of Indian Standards.
19. Standard instrument comprehensive warranty for 5 years + 5 years comprehensive annual maintenance contract (CAMC) after completion of 5 years warranty.
20. Provide point-wise compliance statement indicating compliance of the technical specification vis a vis technical specification offered. Please indicate page number/location of the compliance in the technical brochure.
21. Downtime of equipment > 3 working days after intimation of breakdown to the vendor will accrue a penalty amounting to 0.0005% of the total cost of the equipment per day counted from beyond the first 72-hours of downtime.

From Eye to Insight



To,
All India Institute of Medical Sciences
Ansari Nagar,
New Delhi – 110029

DATED: 19/01/2022

Kind Att: Prof. Tapas Nag

Subject: Quotation for Leica EM UC7 Ultramicrotome

Material	Description	Qty	Net Price Euro
16655816	Dust Cover for control unit	1	
16705805	EM UC7 Basic Instrument	1	
16655850	Touch-sensitive-control unit 10.4" display	1	
16705856	M80 Stereo microscope complete	1	
16700421	Knurled head allen key, 3mm	2	
16701761	Universal specimen holder	2	
16701772	Flat specimen holder, 0-4mm	1	
16705706	Instrument table with top 0.67 x 1.15 m	2	
16705729	Drawer unit with top 0.67 x 0.43 m	1	
16970230	Ultramicrotome accessories	1	
16705525	Grid box for 100 grids, 10 pcs.	1	
16840042	Trufs 6.4 mm, 500 pcs.	1	
16840112	Copper grids 100 mesh, 100 pcs.	1	
16DIA.DU4520	Diamond knife ultra 45° 2.0mm	2	
16DIA.DZ8	Perfect Loop EM set	1	
16DIA.DCO4520	Diamond Knife cryo 45°, dry, 2.0mm	2	
16706031416	Chuck for 3mm carriers	1	
16706031417	Chuck for 6mm carriers	1	
	NET PRICE IN EURO		108,500.00
	Discount		19,000.00

From Eye to Insight

Leica

	NET FINAL CIP PRICE NEW DELHI	89,500.00
--	-------------------------------	-----------

Terms and Conditions

Currency: Euro

Price: CIP New Delhi

Validity of Offer: 90 days from the date of offer.

Delivery: Within 8-10 weeks from the date of receiving advance payment.

Payment Term: 100% advance via Letter of Credit / TT

Country of Origin: As Certified by International Chamber of Commerce

Custom Duty and Clearance: Custom duty payment and clearance to be taken care by the purchaser

Installation and commissioning – Will be done by our trained engineer at your lab

Warranty: 60 Months from the date of installation or 61 months from date of shipment.

Order to be placed on name of:
Leica Mikrosysteme Vertrieb GmbH
Ernst-Leitz-Strasse 17-37
Wetzlar, 35578 Germany

Bank Details:
Name: Deutsche Bank AG, Frankfurt
Bank Number: BLZ 500 700 10
Account Number: 0410 480 00
SWIFT CODE: DEUT DE FF XXX
IBAN: DE84 5007 0010 0041 0480 00

For Leica Microsystems (A Division of DHR Holding India Pvt. Ltd.)



Rohit Kandwal
Sales and Application Manager
Mob: 8800610008

Subhas

TGS

From Eye to Insight



Leica Mikrosysteme Vertrieb GmbH · Ernst-Leitz-Str. 17 - 37 · 35578 Wetzlar, Germany

PROPRIETARY CERTIFICATE

To whomsoever it may concern

This to certify that below mentioned points is unique to Leica EM UC7 Model & is of proprietary nature equipment:

- Vibration decoupled gravity stroke
- Integrated anti-vibration system and grey cast iron base
- Four independent and adjustable LED light sources provide optimal illumination during all portions of the sectioning process
- Unique separate spot light illumination
- Stepping motor 10mm N-S movement of the Knife
- Upgradeable to handle cryosectioning applications with the built-in control of the FC7 cryo chamber, CRION ionizer, as well as for automated serial sectioning processing via the upgrade

These items are solely manufactured by us and to the best of our knowledge no company other than us manufacture these products anywhere else in the world.

We also wish to inform that we are operating in India through our Parent company Danaher called as DHR Holding India Pvt. Ltd.

Thanking you

Yours faithfully,

For and on behalf of Leica Mikrosysteme Vertrieb GmbH

DocuSigned by:
Karim Safa
AC9F1263BBA7400...

Karim Safaee
Manager Commercial Operations

Date : 12/20/2021