

**ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI – 110029**

RE NO 01/PM&R/19-20

Dated 05/04/2019

Call for Rate Enquiry No 192001/19-20

The department of PMR invites sealed quotations by post/per bearer for purchase of Material to be used in P & O workshop Deptt of P.M.R (List attached). Duly sealed quotations should be sent in a sealed cover to reach this office on or before **22/04/2019 up to 12.30 PM**. The quotations shall be opened by the designated purchase committee on **22/04/2019 at 3.30 PM** in Room No. 1, Department of PMR, AIIMS. Participating parties, if interested, may send their representatives to attend the meeting for opening the quotations:

Please ensure the following: -

The rates given in the quotations should be valid for a period of six months.

The prices should be inclusive of all taxes and liabilities till the time of installation and commissioning, if any.


Quotations written in pencil will not be entertained. Any over-writing or erased entries will be disqualified and not considered.

Quotations received later than the above mentioned time shall not be entertained under any circumstances.

Samples would have to be produced whenever asked for. The quality of the samples provided shall be subject to approval by the committee.

Please mark the sealed envelope: **“Quotations for PMR, AIIMS Ref. No 192001/19-20”**

Note: An amount Rs 2,000/- (Rs Two thousand only) in favor of Director AIIMS (demand draft or in cash) may please be deposited with cashier AIIMS towards security (Refundable) otherwise quotation will not be accepted.



(Mr. Guljeet Singh)
Technical Officer (Workshop Supervisor)



(Dr. U. Singh)
Professor and Head

**DEPARTMENT OF PHYSICAL MEDICINE AND REHABILITATION
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI – 110029.**

RE-01/2019-20/PM&R

Dated: 05/04/2019


Pre-Fabricated insoles:


Quantity: thermo-Formable: 400 Pairs, Non Thermo-formable: 100 Pairs.

Specification:

A.	Thermo-formable (400 Pairs) To be used for various foot conditions such as Flat Foot, Planter Fasciitis, Longitudinal High Arch Calcaneum Spur, Metatarsalgia, Excessive Pronation, Calcaneo valgus Varus Problems Closed-Cell EVA material thermo-formable at 110-140 degree centigrade. Shore value between 30-45 The internal Support should have synthetic fibers impregnated with co-polymers. Should be washable. Size: Endurance (38-40, 41-43, 44-46) Fragile (38-40, 41-43, 44-46) Sensible (38-40, 41-43, 44-46) Gliding (38-40, 41-43, 44-46) Rates if different for each size, should be quoted Separately.
B.	Non-Thermo-formable (100 Pairs) (Size-6,7 and 8) Insoles to be used for patients with Diabetes. Open-Cell Material Size-6, 7, and 8 Thickness: 6mm Shore A Hardness Should be able to regains its original thickness on removal of pressure (having Memory) Should have breathable microfiber covering material. Should be able to absorb moisture and remain dry and sweat free. Rates if different for each size, should be quoted Separately.

The above specifications are based on previous purchased (Board Based), generalized in respect of the requirement and not suited to any particular firm/brand.


(Mr. Guljeet Singh)
Technical Officer/Supervisor Workshop


(Dr. U. Singh)
Professor and Head