



COMPREHENSIVE RURAL HEALTH SERVICES PROJECT
CENTRE FOR COMMUNITY MEDICINE
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
BALLABGARH – 121004 (HR)

Ph. 91- 0129-2241362
Fax. 91-0129-2211227

No.F.R.E. 3 /CRHSP-AIIMS/Ballabgarh

Dated the 01.07.2016

To,

Subject: Quotation for purchase of **Infusion Syringe Pump** at CRHSP, Ballabgarh Hospital.

Sir,

Your are requested to please give your quotation for purchase of **Infusion Syringe Pump** at CRHSP, Ballabgarh Hospital. Quotation should reach the office of the **undersigned latest by 15.07.2016**.
Specification is as under:-

- Pumping Mechanism : Lead screw mechanism
- Syringe capacity: 10,20,30, 50/60ml
- Syringe Acceptability: All available standard Indian & imported syringes (Dispovan, B-D, Etc.).

Normal Flow rate range:	Syringe Capacity	Flow Rate
	10ml	0.1 – 300 ml/hr
	20ml	0.1 – 500 ml/hr
	30ml	0.1-700 ml/hr
	50/60 ml	0.1-999.9 ml/hr

Accuracy:	Mechanical accuracy	±1%
	Overall Accuracy*	±2% (Measured time>1hr, min 0.2ml, excluding syringe tolerances.

Occlusion Pressure limits:	Low	0.4+/-0.08kg/cm ² (bar)
	Medium	0.8+/-0.20Kg/cm ² (bar)
	High	1.2+/-0.25 kg/cm ² (bar)

Operating conditions:	Ambient Temperature	10 to 40° C
	Relative Humidity	20 to 90% (on condensation)
	Atmospheric Pressure	70 to 106 kpa

Power:	Power Source	190-280 VAC, 50/60Hz
	Power consumption	12 VA Maximum
	Fuse	2A

Terms and Conditions:-

Comprehensive warranty for two years from the date of purchase.

Standard AIIMS terms and conditions apply.

The responsibility for warranty will be entirely with the local dealer for the full period.

Thanking you,

(Dr. Shashi Kant)
Professor-in-charge