

Specifications for pipette

5 single channel (n=3)

- Adjustable volume
- Short pipetting stroke
- Repeatedly autoclavable at 121°C without disassembly or recalibration
- Separate Tip Ejector with low operating force
- Volume display in 4 digits
- Volume brake to prevent accidental volume change
- Light weight, rugged and resistant to chemical damage
- In-lab calibration adjustments without tools for ISO/GLP compliance
- Accepts all manufacturers' tips
- Accuracy and CE (Single Channel)
 - 0.1-2.5 uL (<2%, <1.5%)
 - 0.5-10ul (<1.5%, <0.8%)
 - 2-20ul (<0.8%, <0.4%)
 - 20-200ul , 100- 1000ul (<0.6%, <0.2%)
- Accuracy and CE (Multiple channel)
 - 0.5-10ul (<1.6%, <1%)
 - 10-100ul (<0.8%, <0.3%)
 - 30-300ul (<0.6%, <0.3%)
- Volume setting should automatically lock the volume
- Should be chemically resistant
- Single-handed operation must be possible
- Should have a 5 year warranty
- Pipettes should be capable of being readjusted and calibrated for use with liquids that have a different density than water
- A service kit and stand should be provided
- CE-IVD compliant
- Shelf mount
- Carousel stand
- 360 degrees rotation of multichannel pipette in both directions to allow pipetting from any angle

Single Channel Pipette

- 0.1-2.5 μ L
- 0.5-10 μ L
- 2-20 μ L
- 20-200 μ L
- 100-1000 μ L

Renu Dhingra
Dr. Renu Dhingra
Professor
Department of Anatomy
A.I.I.M.S., New Delhi-29

A. Shariff
डॉ० ए. शरीफ/Dr. A. SHARIFF
आचार्य/Professor
शरीर रचना विभाग/Deptt. of Anatomy
अ.भा.आ.सं., नई दिल्ली/AIIMS, New Delhi-29



Tony George Jacob
डॉ० टोनी जॉर्ज जेकब
Dr. TONY GEORGE JACOB, MD, Ph.D.
सहायक आचार्य/Assistant Professor
शरीर रचना विभाग/Deptt. of Anatomy
अ भा आ सं , नई दिल्ली-29 /A.I.I.M.S., New Delhi-29

Sarika Ray
19/11/16

Terms & Conditions:

- The quotations should be submitted to the Professor & Head, Department of Anatomy, Room No. 1010, AIIMS, New Delhi-29 and sent by post/per bearer by 7th October, 2016 up to 12.00 P.M. Quotation should be sealed in an envelope and the "name of the item" clearly written on top of the envelop.
- The make of the article offered should invariably be quoted. Quotation should be typed written in ink. All overwriting and erased entries will be deleted from the quotation. The rates should be valid for at least THREE MONTHS.
- VAT/Sales tax or any other kind of tax(s) must be separately mentioned against each item. In case no sales tax is chargeable, prices must be quoted as NET.
- **Warranty/Guarantee:** two years followed by three years CMC comprehensive warranty.
- **Installation:** The firm install the equipment on the site, free of cost
- The firm should be able to give demonstration of the item if required within 5 days of requisition.
- The firm must supply the item within 20 days after confirmation of order.
- **The successful bidder has to submit performances bank guarantee of 10% of the order value valid upto warranty period.**
- The payment will be made electronically viz RTGS/NEFT against delivery and satisfactory installation of the item. The following information should be also mentioned in the quotation.
- Name of the beneficiary:
- Account No. of the beneficiary:
- IFCS code of the bank/branch:
- The quantity can be increased or decreased to any extent depending upon the actual requirement.
- The Professor & Head of the Department reserves the right to cancel/reject full or any part of the rate enquiry which generally do not fulfil the conditions stipulated in the rate enquiry.
- Payment will be made after receiving the items in good condition and satisfactory installation.
- No extra transportation charge will be given to the firm.
- The firm must supply the items within the date mentioned on the supply order copy.

Thanking you

Renu Dhingra
(Dr. Renu Dhingra)

Professor

Computer facility: With request for uploading the same in AIIMS websites for the period of 15 Days.

Dr. Renu Dhingra
Professor
Department of Anatomy
A.I.I.M.S., New Delhi-29