

**ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI-29.
STORES SECTION (DO)**

Ref. No. 34/Stores(DO)/E&M/PAC/2018-19/FSC

Dated-05/02/2019

Sub:- Purchase of "Table Top Refrigerated Centrifuge – 01 No" for the Department of Endo. & Metabolism at AIIMS, New Delhi-110029, on proprietary basis Inviting comments thereon.

The Institute is in the process to purchase **"Table Top Refrigerated Centrifuge – 01 No** at AIIMS, New Delhi from M/s. Thermo Electron LED GmbH, Germany through M/s. Vision Diagnostic (India) Pvt. Ltd., Delhi The PAC Certifications by M/s. Thermo Electron LED GmbH, Germany as well as the user department are attached.

The above documents are being uploaded for open information to submit objections, comments, if any, from any manufacturer regarding proprietary nature of the equipment/item within 15 day from the date of issue/uploading of the notification giving reference No. **34/Stores(DO)/E&M/PAC/2018-19/FSC**. The comments should be received in office of Stores Officer (FSC), Store Section (DO), Animal House Building, Near Biotechnology Building at AIIMS on or before **20/02/2019** upto 12.30 p.m. failing which it will be presumed that any other vendor is having no comment to offer and case will be decided on merits.

Yours faithfully,


SR. STORES OFFICER (DO)

Encl: Related documents enclosed.

9

Dept. of Endo & Metabolism
All India Institute of Medical Science

Dated :- 19th Dec 2018

Sub: Technical Specifications drawn by Technical Committee for the purchase of TABLE TOP REFRIGERATED CENTRIFUGE (2 Nos.)

1) TABLE TOP REFRIGERATED CENTRIFUGE:

ESSENTIAL REQ.

Centrifuge should have capability to fit:

- a] Light weight **Carbon Fiber** swinging bucket rotors to centrifuge at least > 75 blood collection tubes in a single run (5/7 ml yellow cap tubes – 13x100 mm) at 4 degrees.
- b] Light weight **Carbon Fiber** swinging bucket rotors to centrifuge at least 4 Standard micro plates or Deep well+ filter micro plates of height at least > 80 mm in a single run at 4 degrees.
- c] Centrifuge should have a feature to **install and remove rotor with just a push of a button without any tool** for quick and easy change of rotors in less than 5 seconds for different applications

2) ROTORS:

ESSENTIAL REQ –

- a] Swinging bucket rotors must be manufactured from a highly corrosion and fatigue resistant Carbon Fiber material.

b] ESSENTIAL REQ–

Swinging bucket Carbon Fiber Rotor for blood collection tubes

The centrifuge must be supplied with Carbon Fiber Swinging bucket Rotor capable of running at least 76x5 ml/7ml blood collection tubes, 2 sets of adaptors should be supplied. RCF of rotor should be 4600xg.

c] ESSENTIAL REQ–

Carbon Fiber Swinging bucket Rotor for Microplates / Deepwell+ filter plates

The centrifuge must be supplied with a carbon fiber swinging bucket rotor capable of centrifuging at least 4 standard Microplates or Deep well Microplates with filter plate of maximum height at least > 80 mm

3] At least RPM:15000 ; RCF:25500xg

4] Centrifuge Temp range should be -10 °C to +40°C It must display both air/chamber temperature as well as temperature in the sample.

डॉ. निखिल टंडन एम.डी., पी.एच.डी.
Dr. NIKHIL TANDON M.D., Ph.D.
आचार्य एवं अध्यक्ष/Professor & Head
एन्डोक्राइनोलॉजी एवं मेटाबोलिज्म विभाग
Dept. of Endocrinology & Metabolism
आर.आई.एम.एस., नई दिल्ली-110029
All India Institute of Medical Science, New Delhi-110029

(Signature)

(Signature)

डॉ० यशदीप गुप्ता
Dr. YASHDEEP GUPTA
एम.डी., पी.एच.डी. (एन्डो)/M.D., DM, (Endo.)
सहायक आचार्य/Assistant Professor
एन्डोक्राइनोलॉजी एवं मेटाबोलिज्म विभाग
Dept. of Endocrinology & Metabolism
आर.आई.एम.एस., नई दिल्ली-110029
All India Institute of Medical Science, New Delhi-110029

(Signature)
Dr. MANTHA GUPTA Ph.D.
एम.डी., पी.एच.डी. (एन्डो) पी.एच.डी.
सहायक आचार्य/Assistant Professor
एन्डोक्राइनोलॉजी एवं मेटाबोलिज्म विभाग
Dept. of Endocrinology & Metabolism
आर.आई.एम.एस., नई दिल्ली-110029
All India Institute of Medical Science, New Delhi-110029

Dr. MANTHA GUPTA Ph.D.
एम.डी., पी.एच.डी. (एन्डो) पी.एच.डी.
सहायक आचार्य/Assistant Professor
एन्डोक्राइनोलॉजी एवं मेटाबोलिज्म विभाग
Dept. of Endocrinology & Metabolism
आर.आई.एम.एस., नई दिल्ली-110029
All India Institute of Medical Science, New Delhi-110029

- 5] Centrifuge must have Dual Timer mode- At Start and at Speed
- 6] Centrifuge must have an option for automatic lid opening at the end of the run
- 7] Centrifuge must have capability of password protection for the programs and password protection for lid opening.
- 8] Centrifuge must be able to display set parameters together
- 9] Centrifuge must have a minimum of 5 "direct recall" program keys, and capability for up to 99 programs.
- 10] Centrifuge must be CE, CSA and UL certified for safety containment.
- 11] Standard 5 Years Warranty.
- 12] After the warranty is over, CAMC for next 5 years on AIIMS terms and conditions to be submitted by vendor.
- 13] Training - Free installation and onsite Training to user staff.

It is hereby certified that the specification submitted for **TABLE TOP REFRIGERATED CENTRIFUGE** are broad based, general in respect to the requirement and not suit to any particular firm/brand.

Dr. Nikhil Tandon

डॉ. निखिल टण्डन एम.डी., पी.एच.डी.
Dr. NIKHIL TANDON M.D., Ph.D.
आचार्य एवं अध्यक्ष/Professor & Head
अन्तःस्त्राविकी एवं चयापचय विभाग
Deptt. of Endocrinology & Metabolism
अ.भा.आ.सं., नई दिल्ली-110029
A.I.I.M.S., New Delhi-110029

Dr. Mahesh R.

Dr. Anushree Gupta

Dr. Nandita Gupta

डॉ. नन्दिता गुप्ता पी.एच.डी.
Dr. NANDITA GUPTA Ph.D.
PROFESSOR/आचार्य
Dept. of Endocrinology & Metabolism
अन्तःस्त्राविकी एवं चयापचय विभाग
All India Institute of Medical Sciences
अभिलाष भारतीय आयुर्विज्ञान संस्थान
New Delhi-110029/नई दिल्ली-110029

Dr. Yashdeep Gupta

डॉ० यशदीप गुप्ता
Dr. YASHDEEP GUPTA
एम.डी., डी.एच. (एन्डो)/M.D., DM, (Endo.)
सहायक आचार्य/Assistant Professor
अन्तःस्त्राविकी एवं चयापचय विभाग
Deptt. of Endocrinology & Metabolism
अ.भा.आ.सं., नई दिल्ली-110029
A.I.I.M.S., New Delhi-110029

(6)

**ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR : NEW DELHI: 110029**

PROPRIETRY/SPECIFICATION BRAND GOODS CERTIFICATE

Dated - 19th Dec 2018

1. Item/Type/Model No. required along with specification.
2. If the item a spare part attachment or accessory for an existing equipment.
3. Name of the manufacturer/supplier of the item proposed by the indenter.
4. Are they sole manufacturer/sold distributor of the item
5. If there any other item with similar /equivalent specification available in the market to meet the job requirement envisaged. If the answer is yes, why the same can't be procured. Demanding officer should bring out comparative functional advantages /cost effectiveness of the recommended item from these offered by other.
6. What were the efforts made to located alternative source of supply or use other substitutes.
7. Why open/limited tender can't be resorted to, for locating alternative source.
8. Are the proprietary items certifying that the rates are reasonable or not.
9. Any other justification for procuring item from the single source.

Table Top Refrigerated Centrifuge

No, It is a machine.

M/s Thermo Electron LED GmbH Germany. is sole manufacturer and **M/s Vision Diagnostic (India) Pvt. Ltd.** is authorized distributors for the same.

Yes - sole manufacturer


No

M/s Thermo Electron LED GmbH, Germany is the sole manufacturer of this Instrument as per our best knowledge

Because there is no other manufacture as per our best knowledge.

Yes proprietary items & the rates are reasonable.

M/s Thermo Electron LED GmbH, Germany is the sole manufacturer of this Instrument as per our best knowledge



Signature of Indenter
(Demanding Office)

DR. NANDITA GUPTA Ph.D

I certify that the item at Sr. No. 1 above is required to be procured on single tender basics as the source of supply is definitely known / the specified brand proposed was advantages in meeting our functional requirement and limited tender system could be dispensed with as they would serve no useful purpose in this particular case.

(strike out whichever is not applicable)

New Delhi-110029 / New Delhi-110029


डॉ. निखिल टंडन एम.डी., पी.एच.डी.
Dr. NIKHIL TANDON M.D., Ph.D.
आचार्य एवं अध्यक्ष / Professor & Head
अन्तःस्त्राविकी एवं चयापन विभाग
(Counter signed) Endocrinology & Metabolism
(Head of the Department) नई दिल्ली-110029
A.I.I.M.S., New Delhi-110029

7 December 2018

TO WHOM IT MAY CONCERN

This is to inform that the model Thermo Scientific Sorvall legend XTR (Cat No. 75004520) Table Top centrifuge with the following items :



Part Number	Description
75003665	Fiberlite H3-LV Swinging Bucket Rotor
75003180	TX-750 Swinging Bucket Rotor
75003608	4x750 Round Buckets (Set of 4)
75003609	TX-750 Round ClickSeal Biocontainment Lid (Set of 4)
75003723	5/7 mL or 4.5/6 mL Blood Collection Tube

are our proprietary products and being manufactured only at Thermo Electron LED GmbH, Germany.

Some key features of the machine are unique:

- Clickseal biocontainment rotor lid "clicks" when a proper seal has been made, providing assurance that your samples are securely contained, as well as shortens retrieval time with one click opening.
- Auto – Lock III rotor system for push button security and application flexibility
 - Patent No. 8403380, 8678987, 9718067, 9452436, 9339824
- Motorized Lid Latch to reduce closure time and wear
- Fiberlite Carbon Rotor for outstanding ergonomics and performance
 - Patent No. for Swing Rotor with Auto-Lock:
 - 8,147,393
 - 8,211,002
 - 8,282,759
 - 8,323,170
 - 8,328,708

Yours faithfully,

TRUE COPY ATTESTED

O.P. SINGH
Advocate
Regd. No-1610
Notary Public
Ghaziabad

18 DEC 2018



Anita Chng
Commercial Manager, Centrifugation
Asia Pacific
Thermo Fisher Scientific (Pte) Ltd.

Dr. NARAYAN GUPTA, Ph.D.
PROFESSOR/Dean
Dept. of Endocrinology & Metabolism
All India Institute of Medical Sciences
Company Registration No. 28051050327
New Delhi-110028